

Nederlandse norm

# **NEN-EN-IEC 60601-1**

(en)

Medical electrical equipment - Part 1: General requirements for basic safety and essential performance (IEC 60601-1:2005, IDT)

Vervangt NEN 10601-1:1990;  
NEN 10601-1:1990/A2:1995;  
NEN 10601-1:1990/C1:1995;  
NEN 10601-1:1990/A1:1995;  
NEN 10601-1:1990/A13:1996

ICS 11.040.01

oktober 2006

Als Nederlandse norm is aanvaard:

- EN 60601-1:2006,IDT
- IEC 60601-1:2005,IDT

## Normcommissie 301 062 "Medische elektrische toestellen"

---

Apart from exceptions provided by the law, nothing from this publication may be duplicated and/or published by means of photocopy, microfilm, storage in computer files or otherwise, which also applies to full or partial processing, without the written consent of the Netherlands Standardization Institute.

The Netherlands Standardization Institute shall, with the exclusion of any other beneficiary, collect payments owed by third parties for duplication and/or act in and out of law, where this authority is not transferred or falls by right to the Reproduction Rights Foundation.

---

Auteursrecht voorbehouden. Behoudens uitzondering door de wet gesteld mag zonder schriftelijke toestemming van het Nederlands Normalisatie-instituut niets uit deze uitgave worden verveelvoudigd en/of openbaar gemaakt door middel van fotokopie, microfilm, opslag in computerbestanden of anderszins, hetgeen ook van toepassing is op gehele of gedeeltelijke bewerking.

Het Nederlands Normalisatie-instituut is met uitsluiting van ieder ander gerechtigd de door derden verschuldigde vergoedingen voor verveelvoudiging te innen en/of daartoe in en buiten rechte op te treden, voor zover deze bevoegdheid niet is overgedragen c.q. rechtens toekomt aan de Stichting Reprorecht.

---

Although the utmost care has been taken with this publication, errors and omissions cannot be entirely excluded. The Netherlands Standardization Institute and/or the members of the committees therefore accept no liability, not even for direct or indirect damage, occurring due to or in relation with the application of publications issued by the Netherlands Standardization Institute.

---

Hoewel bij deze uitgave de uiterste zorg is nagestreefd, kunnen fouten en onvolledigheden niet geheel worden uitgesloten. Het Nederlands Normalisatie-instituut en/of de leden van de commissies aanvaarden derhalve geen enkele aansprakelijkheid, ook niet voor directe of indirecte schade, ontstaan door of verband houdend met toepassing van door het Nederlands Normalisatie-instituut gepubliceerde uitgaven.

## Nederlands voorwoord

Voor de in deze norm vermelde normatieve verwijzingen bestaan in Nederland de volgende equivalenten:

<u>vermelde norm</u>	<u>Nederlandse norm</u>	<u>titel</u>
IEC 60065:2001	NEN-EN-IEC 60065:2002	Audio-, video- en soortgelijke elektronische toestellen - Veiligheidseisen (en,fr)
IEC 60068-2-2:1974	NEN 10068-2-2:1977	Klimatologische en mechanische beproevingsmethoden voor elektronisch materieel - Deel 2-2: Bestandheid tegen droge warmte (en,fr)
IEC 60068-2-2:1974/A1:1993	NEN-EN-IEC 60068-2-2:1977/A1:1998	Klimatologische en mechanische beproevingsmethoden voor elektrotechnische producten - Deel 2: Beproevingen - Sectie 2: Proef B: Droge warmte (en,fr)
IEC 60068-2-2:1974/A2:1994	NEN 10068-2-2:1977/A2:1994	Klimatologische en mechanische beproevingsmethoden voor elektrotechnische producten - Deel 2: Beproevingen - Sectie 2: Proef B: Droge warmte (en,fr)
IEC 60079-0	NEN-EN-IEC 60079-0	Elektrisch materieel voor plaatsen waar gasontploffingsgevaar kan heersen - Deel 0: Algemene eisen (en,nl)
IEC 60079-2	NEN-EN-IEC 60079-2	Elektrisch materieel voor plaatsen waar gasontploffingsgevaar kan heersen - Deel 2: Inwendige overdruk "p" (en,fr)
IEC 60079-5	NEN-EN 50017	Elektrisch materieel voor plaatsen waar ontploffingsgevaar kan heersen - Zandvulling "q" (en)
IEC 60079-6	NEN-EN 50015	Elektrisch materieel voor plaatsen waar ontploffingsgevaar kan heersen - Olivulling "o" (en)
IEC 60083	-	-
IEC 60085	NEN-EN-IEC 60085	Elektrische isolatiematerialen - Thermische classificatie (en,fr)
IEC 60086-4	NEN-EN-IEC 60086-4	Droge batterijen - Deel 4: Veiligheidsnorm voor lithiumbatterijen (en)
IEC 60112	NEN-EN-IEC 60112	Methode voor de bepaling van de kruipstroomvastheid van vaste isolatiematerialen onder vochtige omstandigheden (en,fr)
IEC 60127-1	NEN-EN-IEC 60127-1	Miniatuursmeltveiligheden - Deel 1: Definities voor miniatuursmeltveiligheden en algemene eisen voor miniatuur(smelt)patronen (en)
IEC 60227-1:1993	-	-
IEC 60227-1:1993/A1:1995	-	-
IEC 60227-1:1993/A2:1998	-	-
IEC 60245-1:2003	-	-
IEC 60252-1	NEN-EN-IEC 60252-1	Wisselstroommotorcondensatoren - Deel 1: Algemeen - Eigenschappen, beproeving en kengegevens - Veiligheidseisen - Handleiding voor installatie en bedrijf (en,fr)
IEC 60320-1	NEN-EN-IEC 60320-1	Toestelstopcontacten voor huishoudelijk en dergelijk algemeen gebruik - Deel 1: Algemene eisen (en,fr)
IEC 60335-1:2001	NEN-EN-IEC 60335-1:2002	Huishoudelijke en soortgelijke elektrische toestellen - Veiligheid - Deel 1: Algemene eisen (en)
IEC 60364-4-41	-	-
IEC 60384-14:2005	NEN-EN-IEC 60384-14:2005	Vaste condensatoren voor elektronische apparatuur - Deel 14: Groepsspecificatie: Vaste condensatoren voor onderdrukking van elektromagnetische storing en verbinding met de voeding (en)
IEC 60417-DB:2002	-	-

IEC 60445	NEN-EN-IEC 60445	Basis- en veiligheidsprincipes voor mens-machine-raakvlak, merken en aanduidingen - Aanduidingen van aansluitklemmen voor elektrisch materieel en voor einden van geleiders met bepaalde functie - Algemene regels voor een alfanumerieke notatie voor klemaanduidingen (en,fr)
IEC 60447	NEN-EN-IEC 60447	Basis- en veiligheidsprincipes voor mens-machine-raakvlak, markering en identificatie - Bedieningsprincipe (en,fr)
IEC 60529:1989	NEN 10529:1992	Beschermingsgraden van omhulsels van elektrisch materieel (IP-codering) (en,fr)
IEC 60529:1989/A1:1999 IEC 60601-1-2	NEN-EN-IEC 60529:1992/A1:2000 NEN-EN-IEC 60601-1-2	Beschermingsgraden van omhulsels (IP-codering) (en,fr) Medische elektrische toestellen - Deel 1-2: Algemene veiligheidseisen - Secundaire norm - Elektromagnetische compatibiliteit - Eisen en beproevingen (en)
IEC 60601-1-3	NEN 10601-1-3	Medische elektrische toestellen - Deel 1: Algemene veiligheidseisen - Sectie 3: Secundaire norm - Algemene eisen voor stralingsbeveiliging in diagnostische röntgentoestellen (en,fr)
IEC 60601-1-6	NEN-EN-IEC 60601-1-6	Medische elektrische toestellen - Deel 1-6: Algemene eisen voor de veiligheid - Secundaire norm - Bruikbaarheid (en,fr)
IEC 60601-1-8	NEN-EN-IEC 60601-1-8	Medische elektrische toestellen - Deel 1-8: Algemene eisen voor de veiligheid - Secundaire norm: Algemene eisen, beproevingen en richtlijnen - Algemene eisen en richtlijnen voor alarmsystemen in medische elektrische toestellen en in medische elektrische systemen (en)
IEC 60664-1:1992	NEN-EN-IEC 60664-1:2003	Coördinatie van isolatie voor inrichtingen binnen laagspanningssystemen - Deel 1: Uitgangspunten, eisen en beproevingen (en,fr)
IEC 60664-1:1992/A1:2000	-	-
IEC 60664-1:1992/A2:2002	-	-
IEC 60695-11-10	NEN-EN-IEC 60695-11-10	Brandbaarheid van elektrotechnische producten - Deel 11-10: Beproevingsvlammen - Beproevingsmethoden voor horizontale en verticale 50 W-vlammen (en,fr)
IEC 60730-1:1999	NEN-EN-IEC 60730-1:2001	Automatische elektrische regelaars voor huishoudelijk en soortgelijk gebruik - Deel 1: Algemene eisen (en,fr)
IEC 60730-1:1999/A1:2003	NEN-EN-IEC 60730-1:1999/A1:2004	Automatische elektrische regelaars voor huishoudelijk en soortgelijk gebruik - Deel 1: Algemene eisen (en,fr)
IEC 60825-1:1993	NEN 10825-1:1994	Veiligheid van laserproducten - Deel 1: Apparatuurclassificatie, eisen en gebruikershandleiding (en,fr)
IEC 60825-1:1993/A1:1997	NEN-EN-IEC 60825-1:1994/A1:2002	Veiligheid van laserproducten - Deel 1: Apparatuurclassificatie, eisen en gebruikershandleiding (en,fr)
IEC 60825-1:1993/A2:2001	NEN-EN-IEC 60825-1:1994/A2:2001	Veiligheid van laserproducten - Deel 1: Apparatuurclassificatie, eisen en gebruikershandleiding (en,fr)
IEC 60851-3:1996	NEN 10851-3:1997	Wikkeldraad - Beproevingsmethoden - Deel 3: Mechanische eigenschappen (en,fr)
IEC 60851-3:1996/A1:1997	NEN-EN-IEC 60851-3:1997/A1:1997	Wikkeldraad - Beproevingsmethoden - Deel 3: Mechanische eigenschappen (en,fr)
IEC 60851-3:1996/A2:2003	NEN-EN-IEC 60851-3:1996/A2:2004	Wikkeldraad - Beproevingsmethoden - Deel 3: Mechanische eigenschappen (en,fr)
IEC 60851-5:1996	NEN 10851-5:1997	Wikkeldraad - Beproevingsmethoden - Deel 5: Elektrische eigenschappen (en,fr)
IEC 60851-5:1996/A1:1997	NEN 10851-5:1997/A1:1997	Wikkeldraad - Beproevingsmethoden - Deel 5: Elektrische eigenschappen (en)

IEC 60851-5:1996/A2:2004 IEC 60851-6:1996	NEN-EN-IEC 60851-5:1997/A2:2004 NEN 10851-6:1997	Wikkeldraad - Beproevingsmethoden - Deel 5: Elektrische eigenschappen (en,fr) Wikkeldraad - Beproevingsmethoden - Deel 6: Thermische eigenschappen (en)
IEC 60851-6:1996/A1:1997 IEC 60878:2003 IEC 60884-1 IEC 60950-1:2001 IEC 61058-1:2000	NEN-EN-IEC 60851-6:1996/A1:1997 - - - NEN-EN-IEC 61058-1:2002	Wikkeldraad - Beproevingsmethoden - Deel 6: Thermische eigenschappen (en,fr) - - - Toestelschakelaars - Deel 1: Algemene eisen (en,fr)
IEC 61058-1:2000/A1:2001 IEC 61558-1:1997 IEC 61558-1:1997/A1:1998 IEC 61558-2-1	- - - NEN-EN-IEC 61558-2-1	- - - Veiligheid van energietransformatoren, voedingseenheden en dergelijke - Deel 2-1: Bijzondere bepalingen voor scheidingstransformatoren voor algemeen gebruik (en,fr)
IEC 61672-1	NEN-EN-IEC 61672-1	Elektro-akoestiek - Geluidniveaumeters - Deel 1: Specificaties (en,fr)
IEC 61672-2	NEN-EN-IEC 61672-2	Elektro-akoestiek - Geluidniveaumeters - Deel 2: Terugkerende evaluatiebeproeving (en,fr)
IEC 61965	NEN-EN-IEC 61965	Mechanische veiligheid van kathodestraalbuizen (en)
ISO 31:series ISO 780	NEN 1226:reeks NEN-EN-ISO 780	Grootheden, SI-eenheden en hun symbolen (nl) Verpakking - Pictogrammen voor behandeling tijdens transport (en)
ISO 1000	NEN 1000	Regels voor het hanteren van het Internationale Stelsel van Eenheden (SI) (nl)
ISO 1853	NEN-ISO 1853	Geleidende en dissiperende gevulcaniseerde of thermoplastische rubber - Meting van de soortelijke weerstand (en)
ISO 2878	NEN-ISO 2878	Gevulcaniseerde rubber - Antistatische en geleidende producten - Bepaling van de elektrische weerstand (en)
ISO 2882 ISO 3746	- NEN-EN-ISO 3746	- Akoestiek - Bepaling van geluidvermogen-niveaus van geluidbronnen - Globale methode met gebruik van een omhullend meetoppervlak boven een reflecterend oppervlak (en)
ISO 3864-1:2002	NEN 3011:2004	Veiligheidskleuren en -tekens in de werkomgeving en in de openbare ruimte (nl)
ISO 5349-1	NEN-EN-ISO 5349-1	Mechanische trillingen - Meting en de beoordeling van blootstelling van het menselijk lichaam aan hand-armtrillingen - Deel 1: Algemene richtlijnen (en)
ISO 7000-DB:2004 ISO 7010:2003 ISO 9614-1	- - NEN-EN-ISO 9614-1	- - Akoestiek - Bepaling van geluidvermogen-niveaus van geluidbronnen via de meting van geluidintensiteiten - Deel 1: Metingen op vaste punten (en)
ISO 10993:series	NEN-EN-ISO 10993:reeks	Biologische evaluatie van medische hulpmiddelen (en)
ISO 11134 ISO 11135 ISO 11137 ISO 13852 ISO 14971:2000	- - - - NEN-EN-ISO 14971:2001	- - - - Medische hulpmiddelen - Toepassing van risicomangement voor medische hulpmiddelen (en)
ISO 15223	-	-

ISO 23529

NEN-ISO 23529

Rubber - Algemene procedures voor het  
voorbereiden en conditioneren van proefstukken  
voor fysieke beproevingsmethoden (en)

EUROPEAN STANDARD

**EN 60601-1**

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2006

ICS 11.040

Supersedes EN 60601-1:1990 + amendments

English version

**Medical electrical equipment**  
**Part 1: General requirements for basic safety**  
**and essential performance**  
(IEC 60601-1:2005)

Appareils électromédicaux  
Partie 1: Exigences générales  
pour la sécurité de base  
et les performances essentielles  
(CEI 60601-1:2005)

Medizinische elektrische Geräte  
Teil 1: Allgemeine Festlegungen  
für die Sicherheit einschließlich  
der wesentlichen Leistungsmerkmale  
(IEC 60601-1:2005)

This European Standard was approved by CENELEC on 2006-09-12. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

**CENELEC**

European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**Central Secretariat: rue de Stassart 35, B - 1050 Brussels**

## Foreword

The text of document 62A/505A/FDIS, future edition 3 of IEC 60601-1, prepared by SC 62A, Common aspects of electrical equipment used in medical practice, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-1 on 2006-09-12.

The following date was fixed:

- latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2007-07-01

This European Standard supersedes EN 60601-1:1990 and its amendments. However, EN 60601-1:1990 remains valid until all the parts 2 that are used in conjunction with it have been withdrawn. No date of withdrawal of conflicting national standards (dow) has therefore been fixed. However, when Part 1 is used for appliances not covered by a part 2, EN 60601-1:1990 is not to be used after 2009-09-12.

This EN 60601-1:2006 has been significantly restructured compared to EN 60601-1:1990. Requirements in the electrical section have been further aligned with those for information technology equipment covered by EN 60950-1 and a requirement for including a RISK MANAGEMENT PROCESS has been added. For an expanded description of this revision, see Clause A.3.

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directives 90/385/EEC and 93/42/EEC. See Annex ZZ.

In this standard the following print types are used:

- requirements and definitions: in roman type;
- *test specifications: in italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS USED THROUGHOUT THIS STANDARD THAT HAVE BEEN DEFINED IN CLAUSE 3 AND ALSO GIVEN IN THE INDEX: IN SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only. In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex G of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk ( \* ) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

Annexes ZA and ZZ have been added by CENELEC.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60073	NOTE	Harmonized as EN 60073:2002 (not modified).
IEC 60086-1	NOTE	Harmonized as EN 60086-1:2001 (not modified).
IEC 60127-6	NOTE	Harmonized as EN 60127-6:1994 (not modified).
IEC 60309-1	NOTE	Harmonized as EN 60309-1:1999 (not modified).
IEC 60317-43	NOTE	Harmonized as EN 60317-43:1997 (not modified).
IEC 60601-1-1	NOTE	Harmonized as EN 60601-1-1:2001 (not modified).
IEC 60601-1-4	NOTE	Harmonized as EN 60601-1-4:1996 + A1:1999 (not modified).
IEC 60601-2-49	NOTE	Harmonized as EN 60601-2-49:2001 (not modified).
IEC 60695-1-1	NOTE	Harmonized as EN 60695-1-1:2000 (not modified).
IEC 60721 series	NOTE	Harmonized in EN 60721 series (not modified).
IEC 60990	NOTE	Harmonized as EN 60990:1999 (not modified).
IEC 61000-4-11	NOTE	Harmonized as EN 61000-4-11:2004 (not modified).
IEC 61010-1	NOTE	Harmonized as EN 61010-1:2001 (not modified).
IEC 61140	NOTE	Harmonized as EN 61140:2002 (not modified).
IEC 62079	NOTE	Harmonized as EN 62079:2001 (not modified).
IEC 62304	NOTE	Harmonized as EN 62304:2006 (not modified).
ISO 407	NOTE	Harmonized as EN ISO 13407:2004 (not modified).
ISO 8041	NOTE	Harmonized as EN ISO 8041:2005 (not modified).
ISO 13485	NOTE	Harmonized as EN ISO 13485:2003 (not modified).

---

### Endorsement notice

The text of the International Standard IEC 60601-1:2005 was approved by CENELEC as a European Standard without any modification.

---

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60065 (mod)	2001	Audio, video and similar electronic apparatus - Safety requirements	EN 60065 + corr. March	2002 2006
IEC 60068-2-2	1974	Environmental testing	EN 60068-2-2 <sup>1)</sup>	1993
A1	1993	Part 2: Tests - Tests B: Dry heat	A1	1993
A2	1994		A2	1994
IEC 60079-0 (mod)	- <sup>2)</sup>	Electrical apparatus for explosive gas atmospheres Part 0: General requirements	EN 60079-0	2006 <sup>3)</sup>
IEC 60079-2	- <sup>2)</sup>	Electrical apparatus for explosive gas atmospheres Part 2: Pressurized enclosures "p"	EN 60079-2 + corr. April	2004 <sup>3)</sup> 2006
IEC 60079-5	- <sup>2)</sup>	Electrical apparatus for explosive gas atmospheres Part 5: Powder filling 'q'	-	-
IEC 60079-6	- <sup>2)</sup>	Electrical apparatus for explosive gas atmospheres Part 6: Oil-immersion "o"	-	-
IEC 60083	- <sup>2)</sup>	Plugs and socket-outlets for domestic and similar general use standardized in member countries of IEC	-	-
IEC 60085	- <sup>2)</sup>	Electrical insulation - Thermal classification	EN 60085	2004 <sup>3)</sup>
IEC 60086-4	- <sup>2)</sup>	Primary batteries Part 4: Safety of lithium batteries	EN 60086-4	2000 <sup>3)</sup>
IEC 60112	- <sup>2)</sup>	Method for the determination of the proof and the comparative tracking indices of solid insulating materials	EN 60112	2003 <sup>3)</sup>

<sup>1)</sup> EN 60068-2-2 includes supplement A:1976 to IEC 60068-2-2.

<sup>2)</sup> Undated reference.

<sup>3)</sup> Valid edition at date of issue.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60127-1	2006	Miniature fuses Part 1: Definitions for miniature fuses and general requirements for miniature fuse-links	EN 60127-1	2006
IEC 60227-1 <sup>4)</sup>	1993	Polyvinyl chloride insulated cables of rated	-	-
A1	1995	voltages up to and including 450/750 V	-	-
A2	1998	Part 1: General requirements	-	-
IEC 60245-1 <sup>5)</sup>	2003	Rubber insulated cables - Rated voltages up to and including 450/750 V Part 1: General requirements	-	-
IEC 60252-1	- <sup>2)</sup>	AC motor capacitors Part 1: General - Performance, testing and rating - Safety requirements - Guide for installation and operation	EN 60252-1	2001 <sup>3)</sup>
IEC 60320-1	- <sup>2)</sup>	Appliance couplers for household and similar general purposes Part 1: General requirements	EN 60320-1	2001 <sup>3)</sup>
IEC 60335-1 (mod)	2001	Household and similar electrical appliances - Safety Part 1: General requirements	EN 60335-1 A11 A12 + corr. July	2002 2004 2006 2006
IEC 60364-4-41 (mod)	2005	Low-voltage electrical installations Part 4-41: Protection for safety - Protection against electric shock	HD 60364-4-41	2006
IEC 60384-14	2005	Fixed capacitors for use in electronic equipment Part 14: Sectional specification - Fixed capacitors for electromagnetic interference suppression and connection to the supply mains	EN 60384-14	2005
IEC 60417	Data base	Graphical symbols for use on equipment	-	-
IEC 60445	- <sup>2)</sup>	Basic and safety principles for man-machine interface, marking and identification - Identification of equipment terminals and of terminations of certain designated conductors, including general rules for an alphanumeric system	EN 60445	2000 <sup>3)</sup>
IEC 60447	- <sup>2)</sup>	Basic and safety principles for man-machine interface, marking and identification - Actuating principles	EN 60447	2004 <sup>3)</sup>

<sup>4)</sup> HD 21.1 S4:2002, *Cables of rated voltages up to and including 450/750 V and having thermoplastic insulation - Part 1: General requirements*, which is related to, but not directly equivalent with, IEC 60227-1, applies instead.

<sup>5)</sup> HD 22.1 S4:2002, *Cables of rated voltages up to and including 450/750 V and having cross-linked insulation - Part 1: General requirements*, which is related to, but not directly equivalent with, IEC 60245-1, applies instead.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60529	1989	Degrees of protection provided by enclosures (IP Code)	EN 60529	1991
A1	1999		+ corr. May A1	1993 2000
IEC 60601-1-2	- <sup>2)</sup>	Medical electrical equipment Part 1-2: General requirements for safety - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2	2001 <sup>3)</sup>
IEC 60601-1-3	- <sup>2)</sup>	Medical electrical equipment Part 1: General requirements for safety - 3. Collateral standard: General requirements for radiation protection in diagnostic X-ray equipment	EN 60601-1-3	1994 <sup>3)</sup>
IEC 60601-1-6	- <sup>2)</sup>	Medical electrical equipment Part 1-6: General requirements for safety - Collateral standard: Usability	EN 60601-1-6	2004 <sup>3)</sup>
IEC 60601-1-8	- <sup>2)</sup>	Medical electrical equipment Part 1-8: General requirements for safety - Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	EN 60601-1-8 + corr. October	2004 <sup>3)</sup> 2006
IEC 60664-1 (mod) + A1 + A2	1992 2000 2002	Insulation coordination for equipment within low-voltage systems Part 1: Principles, requirements and tests	EN 60664-1	2003
IEC 60695-11-10	- <sup>2)</sup>	Fire hazard testing Part 11-10: Test flames - 50 W horizontal and vertical flame test methods	EN 60695-11-10	1999 <sup>3)</sup>
IEC 60730-1 (mod) A1 (mod)	1999 2003	Automatic electrical controls for household and similar use Part 1: General requirements	EN 60730-1 A12 A1 A13 A14	2000 2003 2004 2004 2005
IEC 60825-1 A1 A2	1993 1997 2001	Safety of laser products Part 1: Equipment classification, requirements and user's guide	EN 60825-1 + corr. February A1 A2 + corr. April	1994 1995 2002 2001 2004
IEC 60851-3 A1 A2	1996 1997 2003	Winding wires - Test methods Part 3: Mechanical properties	EN 60851-3 A1 A2	1996 1997 2003
IEC 60851-5 A1 A2	1996 1997 2004	Winding wires - Test methods Part 5: Electrical properties	EN 60851-5 A1 A2	1996 1997 2004
IEC 60851-6 A1	1996 1997	Winding wires - Test methods Part 6: Thermal properties	EN 60851-6 A1	1996 1997
IEC/TR 60878	2003	Graphical symbols for electrical equipment in medical practice	-	-

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60884-1	- <sup>2)</sup>	Plugs and socket-outlets for household and similar purposes Part 1: General requirements	-	-
IEC 60950-1 (mod)	2001	Information technology equipment - Safety Part 1: General requirements	EN 60950-1 <sup>6)</sup> + corr. April A11	2001 2004 2004
IEC 61058-1 (mod) + A1	2000 2001	Switches for appliances Part 1: General requirements	EN 61058-1	2002
IEC 61558-1 (mod) A1	1997 1998	Safety of power transformers, power supply units and similar Part 1: General requirements and tests	EN 61558-1 <sup>7)</sup> + corr. April A1 A11	1997 2003 1998 2003
IEC 61558-2-1	- <sup>2)</sup>	Safety of power transformers, power supply units and similar Part 2-1: Particular requirements for separating transformers for general use	EN 61558-2-1	1997 <sup>3)</sup>
IEC 61672-1	- <sup>2)</sup>	Electroacoustics - Sound level meters Part 1: Specifications	EN 61672-1	2003 <sup>3)</sup>
IEC 61672-2	- <sup>2)</sup>	Electroacoustics - Sound level meters Part 2: Pattern evaluation tests	EN 61672-2	2003 <sup>3)</sup>
IEC 61965	- <sup>2)</sup>	Mechanical safety of cathode ray tubes	EN 61965	2003 <sup>3)</sup>
ISO 31	Series	Quantities and units of space and time	-	-
ISO 780	- <sup>2)</sup>	Packaging - Pictorial marking for handling of goods	EN ISO 780	1999
ISO 1000	- <sup>2)</sup>	SI units and recommendations for the use of their multiples and of certain other units	-	-
ISO 1853	- <sup>2)</sup>	Conducting and dissipative rubbers, vulcanized or thermoplastic - Measurement of resistivity	-	-
ISO 2878	- <sup>2)</sup>	Rubber, vulcanized - Antistatic and conductive products - Determination of electrical resistance	-	-
ISO 2882	- <sup>2)</sup>	Rubber, vulcanized - Antistatic and conductive products for hospital use - Electrical resistance limits	-	-

<sup>6)</sup> EN 60950-1 is superseded by EN 60950-1:2006, which is based on IEC 60950-1:2005, mod.

<sup>7)</sup> EN 61558-1 is superseded by EN 61558-1:2005, which is based on IEC 61558-1:2005.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
ISO 3746	- <sup>2)</sup>	Acoustics - Determination of sound power levels of noise sources using sound pressure - Survey method using an enveloping measurement surface over a reflecting plane	EN ISO 3746	1995
ISO 3864-1	2002	Graphical symbols - Safety colours and safety signs Part 1: Design principles for safety signs in workplaces and public areas	-	-

---

## **Annex ZZ** (informative)

### **Coverage of Essential Requirements of EC Directives**

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directives 90/385/EEC and 93/42/EEC.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directives concerned.

**WARNING:** Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

---

# INTERNATIONAL STANDARD

# IEC 60601-1

Third edition  
2005-12

---

---

## Medical electrical equipment –

### Part 1: General requirements for basic safety and essential performance

*This **English-language** version is derived from the original **bilingual** publication by leaving out all French-language pages. Missing page numbers correspond to the French-language pages.*



Reference number  
IEC 60601-1:2005(E)

## Publication numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series. For example, IEC 34-1 is now referred to as IEC 60034-1.

## Consolidated editions

The IEC is now publishing consolidated versions of its publications. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

## Further information on IEC publications

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology. Information relating to this publication, including its validity, is available in the IEC Catalogue of publications (see below) in addition to new editions, amendments and corrigenda. Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is also available from the following:

- **IEC Web Site** ([www.iec.ch](http://www.iec.ch))

- **Catalogue of IEC publications**

The on-line catalogue on the IEC web site ([www.iec.ch/searchpub](http://www.iec.ch/searchpub)) enables you to search by a variety of criteria including text searches, technical committees and date of publication. On-line information is also available on recently issued publications, withdrawn and replaced publications, as well as corrigenda.

- **IEC Just Published**

This summary of recently issued publications ([www.iec.ch/online\\_news/justpub](http://www.iec.ch/online_news/justpub)) is also available by email. Please contact the Customer Service Centre (see below) for further information.

- **Customer Service Centre**

If you have any questions regarding this publication or need further assistance, please contact the Customer Service Centre:

Email: [custserv@iec.ch](mailto:custserv@iec.ch)  
Tel: +41 22 919 02 11  
Fax: +41 22 919 03 00

# INTERNATIONAL STANDARD

# IEC 60601-1

Third edition  
2005-12

---

---

## Medical electrical equipment –

### Part 1: General requirements for basic safety and essential performance

© IEC 2005 Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembe, PO Box 131, CH-1211 Geneva 20, Switzerland  
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: [inmail@iec.ch](mailto:inmail@iec.ch) Web: [www.iec.ch](http://www.iec.ch)



Commission Electrotechnique Internationale  
International Electrotechnical Commission  
Международная Электротехническая Комиссия

PRICE CODE **XN**

*For price, see current catalogue*

## CONTENTS

FOREWORD.....	21
INTRODUCTION.....	25
1 Scope, object and related standards.....	29
1.1 * Scope .....	29
1.2 Object .....	29
1.3 * Collateral standards.....	29
1.4 * Particular standards.....	31
2 * Normative references.....	31
3 * Terminology and definitions .....	39
4 General requirements.....	79
4.1 * Conditions for application to ME EQUIPMENT or ME SYSTEMS.....	79
4.2 * RISK MANAGEMENT PROCESS for ME EQUIPMENT or ME SYSTEMS .....	79
4.3 * ESSENTIAL PERFORMANCE .....	81
4.4 * EXPECTED SERVICE LIFE .....	81
4.5 * Equivalent safety for ME EQUIPMENT or ME SYSTEMS .....	83
4.6 * ME EQUIPMENT or ME SYSTEM parts that contact the PATIENT .....	83
4.7 * SINGLE FAULT CONDITION for ME EQUIPMENT.....	83
4.8 Components of ME EQUIPMENT .....	85
4.9 * Use of COMPONENTS WITH HIGH-INTEGRITY CHARACTERISTICS in ME EQUIPMENT .....	85
4.10 * Power supply .....	87
4.11 Power input.....	89
5 * General requirements for testing ME EQUIPMENT .....	91
5.1 * TYPE TESTS.....	91
5.2 * Number of samples .....	91
5.3 Ambient temperature, humidity, atmospheric pressure.....	91
5.4 Other conditions .....	91
5.5 Supply voltages, type of current, nature of supply, frequency .....	93
5.6 Repairs and modifications .....	93
5.7 * Humidity preconditioning treatment .....	93
5.8 Sequence of tests .....	95
5.9 * Determination of APPLIED PARTS and ACCESSIBLE PARTS .....	95
6 * Classification of ME EQUIPMENT and ME SYSTEMS.....	99
6.1 General .....	99
6.2 * Protection against electric shock.....	99
6.3 * Protection against harmful ingress of water or particulate matter .....	101
6.4 Method(s) of sterilization .....	101
6.5 Suitability for use in an OXYGEN RICH ENVIRONMENT .....	101
6.6 * Mode of operation .....	101

7	ME EQUIPMENT identification, marking and documents .....	101
7.1	General .....	101
7.2	Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts .....	105
7.3	Marking on the inside of ME EQUIPMENT or ME EQUIPMENT parts .....	113
7.4	Marking of controls and instruments .....	117
7.5	Safety signs .....	119
7.6	Symbols .....	121
7.7	Colours of the insulation of conductors .....	121
7.8	* Indicator lights and controls .....	123
7.9	ACCOMPANYING DOCUMENTS .....	123
8	* Protection against electrical HAZARDS from ME EQUIPMENT .....	135
8.1	Fundamental rule of protection against electric shock.....	135
8.2	Requirements related to power sources.....	137
8.3	Classification of APPLIED PARTS .....	137
8.4	Limitation of voltage, current or energy.....	139
8.5	Separation of parts.....	145
8.6	* Protective earthing, functional earthing and potential equalization of ME EQUIPMENT .....	161
8.7	LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS .....	167
8.8	Insulation .....	201
8.9	* CREEPAGE DISTANCES and AIR CLEARANCES.....	213
8.10	Components and wiring .....	243
8.11	MAINS PARTS, components and layout .....	247
9	* Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	259
9.1	MECHANICAL HAZARDS of ME EQUIPMENT .....	259
9.2	* HAZARDS associated with moving parts.....	261
9.3	* HAZARD associated with surfaces, corners and edges.....	271
9.4	* Instability HAZARDS .....	271
9.5	* Expelled parts HAZARD .....	281
9.6	Acoustic energy (including infra- and ultrasound) and vibration .....	281
9.7	* Pressure vessels and parts subject to pneumatic and hydraulic pressure.....	285
9.8	* HAZARDS associated with support systems .....	291
10	* Protection against unwanted and excessive radiation HAZARDS .....	301
10.1	X-Radiation .....	301
10.2	Alpha, beta, gamma, neutron and other particle radiation .....	303
10.3	Microwave radiation .....	303
10.4	* Lasers and light emitting diodes (LEDs) .....	303
10.5	Other visible electromagnetic radiation .....	303
10.6	Infrared radiation .....	305
10.7	Ultraviolet radiation .....	305
11	* Protection against excessive temperatures and other HAZARDS.....	305
11.1	* Excessive temperatures in ME EQUIPMENT.....	305
11.2	* Fire prevention.....	313
11.3	* Constructional requirements for fire ENCLOSURES of ME EQUIPMENT .....	323

11.4	* ME EQUIPMENT and ME SYSTEMS intended for use with flammable anaesthetics .....	329
11.5	* ME EQUIPMENT and ME SYSTEMS intended for use in conjunction with flammable agents .....	329
11.6	Overflow, spillage, leakage, ingress of water or particulate matter, cleaning, disinfection, sterilization and compatibility with substances used with the ME EQUIPMENT .....	329
11.7	Biocompatibility of ME EQUIPMENT and ME SYSTEMS .....	333
11.8	* Interruption of the power supply / SUPPLY MAINS to ME EQUIPMENT .....	333
12	* Accuracy of controls and instruments and protection against hazardous outputs .....	333
12.1	Accuracy of controls and instruments .....	333
12.2	USABILITY .....	333
12.3	Alarm systems .....	333
12.4	Protection against hazardous output .....	333
13	* HAZARDOUS SITUATIONS and fault conditions .....	337
13.1	Specific HAZARDOUS SITUATIONS .....	337
13.2	SINGLE FAULT CONDITIONS .....	339
14	* PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) .....	351
14.1	* General .....	351
14.2	* Documentation .....	351
14.3	* RISK MANAGEMENT plan .....	353
14.4	* PEMS DEVELOPMENT LIFE-CYCLE .....	353
14.5	* Problem resolution .....	353
14.6	RISK MANAGEMENT PROCESS .....	353
14.7	* Requirement specification .....	355
14.8	* Architecture .....	355
14.9	* Design and implementation .....	357
14.10	* VERIFICATION .....	357
14.11	* PEMS VALIDATION .....	357
14.12	* Modification .....	359
14.13	* Connection of PEMS by NETWORK/DATA COUPLING to other equipment .....	359
15	Construction of ME EQUIPMENT .....	359
15.1	* Arrangements of controls and indicators of ME EQUIPMENT .....	359
15.2	* Serviceability .....	359
15.3	Mechanical strength .....	361
15.4	ME EQUIPMENT components and general assembly .....	369
15.5	* MAINS SUPPLY TRANSFORMERS of ME EQUIPMENT and transformers providing separation in accordance with 8.5 .....	379
16	* ME SYSTEMS .....	387
16.1	* General requirements for the ME SYSTEMS .....	387
16.2	* ACCOMPANYING DOCUMENTS of an ME SYSTEM .....	389
16.3	* Power supply .....	391
16.4	ENCLOSURES .....	391
16.5	* SEPARATION DEVICES .....	391
16.6	* LEAKAGE CURRENTS .....	393
16.7	* Protection against MECHANICAL HAZARDS .....	395

16.8	Interruption of the power supply to parts of an ME SYSTEM .....	395
16.9	ME SYSTEM connections and wiring .....	395
17	* Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	399
Annex A (informative)	General guidance and rationale.....	401
Annex B (informative)	Sequence of testing .....	613
Annex C (informative)	Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	621
Annex D (informative)	Symbols on marking.....	629
Annex E (informative)	Examples of the connection of the measuring device (MD) for measurement of the PATIENT LEAKAGE CURRENT and PATIENT AUXILIARY CURRENT .....	647
Annex F (informative)	Suitable measuring supply circuits.....	651
Annex G (normative)	Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	657
Annex H (informative)	PEMS structure, PEMS DEVELOPMENT LIFE-CYCLE and documentation .....	687
Annex I (informative)	ME SYSTEMS aspects.....	713
Annex J (informative)	Survey of insulation paths.....	725
Annex K (informative)	Simplified PATIENT LEAKAGE CURRENT diagrams .....	731
Annex L (normative)	Insulated winding wires for use without interleaved insulation.....	737
Bibliography.....		743
INDEX .....		749
INDEX OF ABBREVIATIONS AND ACRONYMS .....		775
Figure 1 – Detachable mains connection.....		43
Figure 2 – Example of the defined terminals and conductors.....		45
Figure 3 – Example of a CLASS I ME EQUIPMENT.....		47
Figure 4 – Example of a metal-enclosed CLASS II ME EQUIPMENT .....		47
Figure 5 – Schematic flow chart for component qualification .....		87
Figure 6 – Standard test finger.....		97
Figure 7 – Test hook.....		99
Figure 8 – Test pin.....		141
Figure 9 – Application of test voltage to bridged PATIENT CONNECTIONS for DEFIBRILLATION-PROOF APPLIED PARTS.....		155
Figure 10 – Application of test voltage to individual PATIENT CONNECTIONS for DEFIBRILLATION-PROOF APPLIED PARTS.....		159
Figure 11 – Application of test voltage to test the delivered defibrillation energy .....		161

Figure 12 – Example of a measuring device and its frequency characteristics.....	169
Figure 13 – Measuring circuit for the EARTH LEAKAGE CURRENT of CLASS I ME equipment, with or without APPLIED PART .....	175
Figure 14 – Measuring circuit for the TOUCH CURRENT.....	177
Figure 15 – Measuring circuit for the PATIENT LEAKAGE CURRENT from the PATIENT CONNECTION to earth.....	179
Figure 16 – Measuring circuit for the PATIENT LEAKAGE CURRENT via the PATIENT CONNECTION(S) of an F-TYPE APPLIED PART to earth caused by an external voltage on the PATIENT CONNECTION(S) .....	181
Figure 17 – Measuring circuit for the PATIENT LEAKAGE CURRENT from PATIENT CONNECTION(S) to earth caused by an external voltage on a SIGNAL INPUT/OUTPUT PART .....	183
Figure 18 – Measuring circuit for the PATIENT LEAKAGE CURRENT from PATIENT CONNECTION(S) to earth caused by an external voltage on a metal ACCESSIBLE PART that is not PROTECTIVELY EARTHED .....	185
Figure 19 – Measuring circuit for the PATIENT AUXILIARY CURRENT .....	187
Figure 20 – Measuring circuit for the total PATIENT LEAKAGE CURRENT with all PATIENT CONNECTIONS of all APPLIED PARTS of the same type (TYPE B APPLIED PARTS, TYPE BF APPLIED PARTS or TYPE CF APPLIED PARTS) connected together.....	189
Figure 21 – Ball-pressure test apparatus .....	213
Figure 22 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 1 .....	239
Figure 23 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 2 .....	239
Figure 24 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 3 .....	239
Figure 25 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 4 .....	239
Figure 26 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 5 .....	239
Figure 27 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 6 .....	241
Figure 28 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 7 .....	241
Figure 29 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 8 .....	241
Figure 30 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 9 .....	241
Figure 31 – CREEPAGE DISTANCE and AIR CLEARANCE – Example 10 .....	243
Figure 32 – Ratio between HYDRAULIC TEST PRESSURE and MAXIMUM PERMISSIBLE WORKING PRESSURE .....	289
Figure 33 – Human body test mass.....	299
Figure 34 – Spark ignition test apparatus.....	317
Figure 35 – Maximum allowable current $I$ as a function of the maximum allowable voltage $U$ measured in a purely resistive circuit in an OXYGEN RICH ENVIRONMENT .....	317
Figure 36 – Maximum allowable voltage $U$ as a function of the capacitance $C$ measured in a capacitive circuit used in an OXYGEN RICH ENVIRONMENT .....	319
Figure 37 – Maximum allowable current $I$ as a function of the inductance $L$ measured in an inductive circuit in an OXYGEN RICH ENVIRONMENT .....	319
Figure 38 – Baffle .....	327
Figure 39 – Area of the bottom of an ENCLOSURE as specified in 11.3 b) 1) .....	327
Figure A.1 – Identification of ME EQUIPMENT, APPLIED PARTS and PATIENT CONNECTIONS in an ECG monitor .....	413

Figure A.2 – Example of the insulation of an F-TYPE APPLIED PART with the insulation incorporated in the ME EQUIPMENT .....	415
Figure A.3 – Identification of ME EQUIPMENT, APPLIED PARTS and PATIENT CONNECTIONS in a PATIENT monitor with invasive pressure monitoring facility .....	417
Figure A.4 – Identification of ME EQUIPMENT, APPLIED PARTS and PATIENT CONNECTIONS in a multifunction PATIENT monitor with invasive pressure monitoring facilities .....	419
Figure A.5 – Identification of APPLIED PARTS and PATIENT CONNECTIONS in an X-ray ME SYSTEM .....	421
Figure A.6 – Identification of ME EQUIPMENT, APPLIED PARTS and PATIENT CONNECTIONS in a transcutaneous electronic nerve stimulator (TENS) intended to be worn on the PATIENT'S belt and connected to electrodes applied to the PATIENT'S upper arm.....	421
Figure A.7 – Identification of ME EQUIPMENT or ME SYSTEM, APPLIED PARTS and PATIENT CONNECTIONS in a personal computer with an ECG module .....	423
Figure A.8 – Pictorial representation of the relationship of HAZARD, sequence of events, HAZARDOUS SITUATION and HARM .....	429
Figure A.9 – Example of PATIENT ENVIRONMENT.....	441
Figure A.10 – Floating circuit .....	469
Figure A.11 – Interruption of a power-carrying conductor between ME EQUIPMENT parts in separate ENCLOSURES.....	475
Figure A.12 – Identification of MEANS OF PATIENT PROTECTION and MEANS OF OPERATOR PROTECTION.....	483
Figure A.13 – Allowable protective earth impedance where the fault current is limited .....	497
Figure A.14 – Probability of ventricular fibrillation .....	509
Figure A.15 – Example of a measuring circuit for the PATIENT LEAKAGE CURRENT from a PATIENT CONNECTION to earth for ME EQUIPMENT with multiple PATIENT CONNECTIONS .....	519
Figure A.16 – Instability test conditions.....	543
Figure A.17 – Example of determining TENSILE SAFETY FACTOR using Table 21 .....	555
Figure A.18 – Example of determining design and test loads .....	557
Figure A.19 – Example of human body mass distribution .....	557
Figure E.1 – TYPE B APPLIED PART.....	647
Figure E.2 – TYPE BF APPLIED PART .....	647
Figure E.3 – TYPE CF APPLIED PART .....	649
Figure E.4 – PATIENT AUXILIARY CURRENT .....	649
Figure E.5 – Loading of the PATIENT CONNECTIONS if specified by the MANUFACTURER .....	649
Figure F.1 – Measuring supply circuit with one side of the SUPPLY MAINS at approximately earth potential.....	651
Figure F.2 – Measuring supply circuit with SUPPLY MAINS approximately symmetrical to earth potential.....	651
Figure F.3 – Measuring supply circuit for polyphase ME EQUIPMENT specified for connection to a polyphase SUPPLY MAINS.....	653
Figure F.4 – Measuring supply circuit for single-phase ME EQUIPMENT specified for connection to a polyphase SUPPLY MAINS.....	653

Figure F.5 – Measuring supply circuit for ME EQUIPMENT having a separate power supply unit or intended to receive its power from another equipment in an ME SYSTEM.....	655
Figure G.1– Maximum allowable current $I_{ZR}$ as a function of the maximum allowable voltage $U_{ZR}$ measured in a purely resistive circuit with the most flammable mixture of ether vapour with air .....	669
Figure G.2 – Maximum allowable voltage $U_{ZC}$ as a function of the capacitance $C_{max}$ measured in a capacitive circuit with the most flammable mixture of ether vapour with air ..	671
Figure G.3 – Maximum allowable current $I_{ZL}$ as a function of the inductance $L_{max}$ measured in an inductive circuit with the most flammable mixture of ether vapour with air ..	671
Figure G.4 – Maximum allowable current $I_{ZR}$ as a function of the maximum allowable voltage $U_{ZR}$ measured in a purely resistive circuit with the most flammable mixture of ether vapour with oxygen .....	679
Figure G.5 – Maximum allowable voltage $U_{ZC}$ as a function of the capacitance $C_{max}$ measured in a capacitive circuit with the most flammable mixture of ether vapour with oxygen.....	681
Figure G.6 – Maximum allowable current $I_{ZL}$ as a function of the inductance $L_{max}$ measured in an inductive circuit with the most flammable mixture of ether vapour with oxygen.....	681
Figure G.7 – Test apparatus .....	685
Figure H.1 – Examples of PEMS/ PESS structures .....	689
Figure H.2 – A PEMS DEVELOPMENT LIFE-CYCLE model .....	691
Figure H.3 – PEMS documentation requirements from Clause 14 and ISO 14971:2000 .....	699
Figure H.4 – Example of potential parameters required to be specified for NETWORK/DATA COUPLING .....	711
Figure I.1 – Example of the construction of a MULTIPLE SOCKET-OUTLET (MSO).....	721
Figure I.2 – Examples of application of MULTIPLE SOCKET-OUTLETS (MSO) .....	723
Figure J.1 – Insulation example 1 .....	725
Figure J.2 – Insulation example 2 .....	725
Figure J.3 – Insulation example 3 .....	725
Figure J.4 – Insulation example 4 .....	727
Figure J.5 – Insulation example 5 .....	727
Figure J.6 – Insulation example 6 .....	727
Figure J.7 – Insulation example 7 .....	729
Figure K.1 – ME EQUIPMENT with an ENCLOSURE made of insulating material.....	731
Figure K.2 – ME EQUIPMENT with an F-TYPE APPLIED PART.....	731
Figure K.3 – ME EQUIPMENT with an APPLIED PART and a SIGNAL INPUT/OUTPUT PART .....	733
Figure K.4 – ME EQUIPMENT with a PATIENT CONNECTION of a TYPE B APPLIED PART that is not PROTECTIVELY EARTHED .....	733
Figure K.5 – ME EQUIPMENT with a PATIENT CONNECTION of a TYPE BF APPLIED PART that is not PROTECTIVELY EARTHED .....	735

Table 1 – Units outside the SI units system that may be used on ME EQUIPMENT .....	119
Table 2 – Colours of indicator lights and their meaning for ME EQUIPMENT .....	123
Table 3 – * Allowable values of PATIENT LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS under NORMAL CONDITION and SINGLE FAULT CONDITION.....	171
Table 4 – * Allowable values of PATIENT LEAKAGE CURRENTS under the special test conditions identified in 8.7.4.7 .....	173
Table 5 – Legends of symbols for Figure 9 to Figure 11, Figure 13 to Figure 20, Figure A.15, Annex E and Annex F .....	191
Table 6 – Test voltages for solid insulation forming a MEANS OF PROTECTION .....	207
Table 7 – Test voltages for MEANS OF OPERATOR PROTECTION.....	209
Table 8 – Multiplication factors for AIR CLEARANCES for altitudes up to 5 000 m.....	215
Table 9 – Material group classification .....	217
Table 10 – MAINS TRANSIENT VOLTAGE.....	219
Table 11 – Minimum CREEPAGE DISTANCES and AIR CLEARANCES between parts of opposite polarity of the MAINS PART .....	223
Table 12 – Minimum CREEPAGE DISTANCES and AIR CLEARANCES providing MEANS OF PATIENT PROTECTION .....	225
Table 13 – Minimum AIR CLEARANCES providing MEANS OF OPERATOR PROTECTION from the MAINS PART .....	227
Table 14 – Additional AIR CLEARANCES for insulation in MAINS PARTS with PEAK WORKING VOLTAGES exceeding the peak value of the NOMINAL MAINS VOLTAGE <sup>a</sup> .....	229
Table 15 – Minimum AIR CLEARANCES for MEANS OF OPERATOR PROTECTION in SECONDARY CIRCUITS.....	231
Table 16 – Minimum CREEPAGE DISTANCES providing MEANS OF OPERATOR PROTECTION .....	233
Table 17 – NOMINAL cross-sectional area of conductors of a POWER SUPPLY CORD .....	251
Table 18 – Testing of cord anchorages .....	253
Table 19 – MECHANICAL HAZARDS covered by this clause .....	261
Table 20 – Acceptable gaps.....	265
Table 21 – Determination of TENSILE SAFETY FACTOR .....	293
Table 22 – Allowable maximum temperatures of parts.....	305
Table 23 – Allowable maximum temperatures for ME EQUIPMENT parts that are likely to be touched.....	307
Table 24 – Allowable maximum temperatures for skin contact with ME EQUIPMENT APPLIED PARTS .....	307
Table 25 – Acceptable perforation of the bottom of an ENCLOSURE .....	325
Table 26 – * Temperature limits of motor windings.....	345
Table 27 – Maximum motor winding steady-state temperature .....	349
Table 28 – Mechanical strength test applicability .....	361
Table 29 – Drop height .....	365
Table 30 – Test torques for rotating controls.....	377

Table 31 – Maximum allowable temperatures of transformer windings under overload and short-circuit conditions at 25 °C (± 5 °C) ambient temperature .....	381
Table 32 – Test current for transformers .....	383
Table A.1 – Values of AIR CLEARANCE and CREEPAGE DISTANCE derived from Table 7 of IEC 61010-1:2001 and Table 12 .....	525
Table A.2 – CREEPAGE DISTANCES to avoid failure due to tracking from IEC 60664-1 .....	527
Table A.3 – Instability test conditions .....	543
Table A.4 – Allowable time exposure for level of acceleration .....	547
Table A.5 – Guidance on surface temperatures for ME EQUIPMENT that creates low temperatures (cools) for therapeutic purposes or as part of its operation .....	565
Table C.1– Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts .....	621
Table C.2 – Marking on the inside of ME EQUIPMENT, ME SYSTEMS or their parts .....	623
Table C.3 – Marking of controls and instruments.....	623
Table C.4 – ACCOMPANYING DOCUMENTS, general.....	625
Table C.5 – ACCOMPANYING DOCUMENTS, instructions for use.....	625
Table D.1 – General symbols.....	631
Table D.2 – Safety signs .....	641
Table D.3 – General codes .....	645
Table G.1 – Gas-tightness of cord inlets .....	675
Table H.1 – NETWORK/DATA COUPLING classification .....	707
Table I.1 – Some examples of ME SYSTEMS for illustration .....	717
Table L.1– Mandrel diameter .....	739
Table L.2 – Oven temperature .....	739

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**MEDICAL ELECTRICAL EQUIPMENT –****Part 1: General requirements for basic safety  
and essential performance**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-1 has been prepared by subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 1988, its Amendment 1 (1991) and Amendment 2 (1995). This edition constitutes a technical revision. This edition has been significantly restructured. Requirements in the electrical section have been further aligned with those for information technology equipment covered by IEC 60950-1 and a requirement for including a RISK MANAGEMENT PROCESS has been added. For an expanded description of this revision, see Clause A.3.

The text of this standard is based on the following documents:

FDIS	Report on voting
62A/505A/FDIS	62A/512/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard the following print types are used:

- Requirements and definitions: in roman type.
- *Test specifications: in italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS USED THROUGHOUT THIS STANDARD THAT HAVE BEEN DEFINED IN CLAUSE 3 AND ALSO GIVEN IN THE INDEX: IN SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex G of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk ( \* ) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under “<http://webstore.iec.ch>” in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

## INTRODUCTION

In 1976, IEC subcommittee 62A published the first edition of IEC/TR 60513, *Basic aspects of the safety philosophy for electrical equipment used in medical practice*. The first edition of IEC/TR 60513 provided the basis for developing:

- the first edition of IEC 60601-1 (the parent safety standard for MEDICAL ELECTRICAL EQUIPMENT);
- the IEC 60601-1-xx series of collateral standards for MEDICAL ELECTRICAL EQUIPMENT;
- the IEC 60601-2-xx series of particular standards for particular types of MEDICAL ELECTRICAL EQUIPMENT; and
- the IEC 60601-3-xx series of performance standards for particular types of MEDICAL ELECTRICAL EQUIPMENT.

Aware of the need and the urgency for a standard covering electrical equipment used in medical practice, the majority of National Committees voted in 1977 in favour of the first edition of IEC 60601-1, based on a draft that at the time represented a first approach to the problem. The extent of the scope, the complexity of the equipment concerned, and the specific nature of some of the protective measures and the corresponding tests for verifying them, required years of effort in order to prepare this first standard, which can now be said to have served as a universal reference since its publication.

However, the frequent application of the first edition revealed room for improvement. These improvements were all the more desirable in view of the considerable success that this standard has enjoyed since its publication.

The careful work of revision subsequently undertaken and continued over a number of years resulted in the publication of the second edition in 1988. This edition incorporated all the improvements that could be reasonably expected up to that time. Further developments remained under constant study. The second edition was amended in 1991 and then again in 1995.

The original IEC approach was to prepare separate BASIC SAFETY and performance standards for MEDICAL ELECTRICAL EQUIPMENT. This was a natural extension of the historical approach taken at the national and international level with other electrical equipment standards (e.g. those for domestic equipment), where BASIC SAFETY is regulated through mandatory standards but other performance specifications are regulated by market pressure. In this context, it has been said that, "The ability of an electric kettle to boil water is not critical to its safe use!"

It is now recognized that this is not the situation with many items of MEDICAL ELECTRICAL EQUIPMENT, and RESPONSIBLE ORGANIZATIONS have to depend on standards to ensure ESSENTIAL PERFORMANCE as well as BASIC SAFETY. Such areas include the accuracy with which the equipment controls the delivery of energy or therapeutic substances to the PATIENT, or processes and displays physiological data that will affect PATIENT management.

This recognition means that separating BASIC SAFETY and performance is somewhat inappropriate in addressing the HAZARDS that result from inadequate design of MEDICAL ELECTRICAL EQUIPMENT. Many particular standards in the IEC 60601-2-xx series address a range of ESSENTIAL PERFORMANCE requirements that cannot be directly evaluated by the RESPONSIBLE ORGANIZATION without applying such standards. (However, the current IEC 60601 series includes fewer requirements for ESSENTIAL PERFORMANCE than for BASIC SAFETY).

In anticipation of a third edition of IEC 60601-1, IEC subcommittee 62A prepared a second edition of IEC/TR 60513 [12]<sup>1)</sup> in 1994. It was intended that the second edition of IEC/TR 60513 would provide guidance for developing this edition of IEC 60601-1, and for the further development of the IEC 60601-1-xx and IEC 60601-2-xx series.

In order to achieve consistency in international standards, address present expectations in the health care community and align with developments in IEC 60601-2-xx, the second edition of IEC/TR 60513 includes two major new principles:

- the first change is that the concept of “SAFETY” has been broadened from the BASIC SAFETY considerations in the first and second editions of IEC 60601-1 to include ESSENTIAL PERFORMANCE matters, (e.g. the accuracy of physiological monitoring equipment). Application of this principle leads to the change of the title of this publication from “Medical electrical equipment, Part 1: General requirements for safety” in the second edition, to “Medical electrical equipment, Part 1: General requirements for basic safety and essential performance”;
- the second change is that, in specifying minimum safety requirements, provision is made for assessing the adequacy of the design PROCESS when this is the only practical method of assessing the safety of certain technologies such as programmable electronic systems. Application of this principle is one of the factors leading to introduction of a general requirement to carry out a RISK MANAGEMENT PROCESS. In parallel with the development of the third edition of IEC 60601-1, a joint project with ISO/TC 210 resulted in the publication of a general standard for RISK MANAGEMENT of medical devices. Compliance with this edition of IEC 60601-1 requires that the MANUFACTURER have a RISK MANAGEMENT PROCESS complying with ISO 14971 in place (see 4.2).

This standard contains requirements concerning BASIC SAFETY and ESSENTIAL PERFORMANCE that are generally applicable to MEDICAL ELECTRICAL EQUIPMENT. For certain types of MEDICAL ELECTRICAL EQUIPMENT, these requirements are either supplemented or modified by the special requirements of a collateral or particular standard. Where particular standards exist, this standard should not be used alone.

---

1) Figures in square brackets refer to the Bibliography.

## Bestel uw norm via dit formulier

Ja, ik bestel:

... ex. NEN-EN-IEC 60601-1:2006 (en) Medische elektrische toestellen - Deel 1: Algemene eisen voor basisveiligheid en essentiële prestaties 210,66

Wilt u regelmatig per e-mail op de hoogte worden gehouden van normen en normalisatie? Geef u dan op voor een van onze gratis e-mailnieuwsbrieven\*:

- |   |   |
|---|---|
| <input type="radio"/> Agro-FoodMail (1 x per kwartaal)        | <input type="radio"/> InstallatieMail (1 x per maand)           |
| <input type="radio"/> ArboMail (1 x per maand)                | <input type="radio"/> KAM-Mail (1 x per twee maanden)           |
| <input type="radio"/> BouwMail (1 x per maand)                | <input type="radio"/> MachinebouwMail (1 x per twee maanden)    |
| <input type="radio"/> ChemieMail (1 x per kwartaal)           | <input type="radio"/> MediscopeMail (1 x per twee maanden)      |
| <input type="radio"/> ConsumentenzakenMail (1 x per kwartaal) | <input type="radio"/> MilieuMail (verschijnt 1 x per kwartaal)  |
| <input type="radio"/> ElektroMail (1 x per maand)             | <input type="radio"/> NormmanagementMail (1 x per twee maanden) |
| <input type="radio"/> ICT-Mail (1 x per kwartaal)             |   |

Zie ook <http://nieuwsbrieven.nen.nl>

\* (vul bij de adresgegevens uw e-mailadres in).

Stuur dit formulier ingevuld retour aan: NEN-Uitgeverij, t.a.v. afdeling Marketing, Antwoordnummer 10214, 2600 WB Delft (een postzegel is niet nodig). U kunt ook faxen naar (015) 2 690 186.

Bedrijf / Instelling: \_\_\_\_\_

T.a.v.: \_\_\_\_\_

 dhr. /  mw.

Functie: \_\_\_\_\_

E-mail\*: \_\_\_\_\_

Klantnummer (indien bekend) \_\_\_\_\_

Postbus / Adres: \_\_\_\_\_

Postcode &amp; Plaats: \_\_\_\_\_

Telefoon: \_\_\_\_\_

Fax: \_\_\_\_\_

Factuuradres (indien dit afwijkt van bovenstaand adres): \_\_\_\_\_

Datum: \_\_\_\_\_

Handtekening: \_\_\_\_\_

Mprev \_\_\_\_\_

- De prijzen zijn geldig tot 31 december 2006, tenzij anders aangegeven bij een actie op het bestelformulier.
- Bestellingen van bedrijven die buiten Nederland gevestigd zijn krijgen een toeslag berekend.
- De jaarkosten zijn per abonnement, exclusief btw, exclusief verzend- en administratiekosten, in euro's en onder voorbehoud.
- Heeft u al een abonnement op een van de genoemde normen, dan hoeft u niets te bestellen, u ontvangt de nieuwe norm(en) automatisch. Meer informatie over abonnementen: NEN-Abonnementen: (015) 2 690 391, optie 3 (van 09.00 - 13.00 uur).
- U kunt onze algemene voorwaarden terugvinden op <http://leveringsvoorwaarden.nen.nl>.
- Wijzigingen en typefouten in teksten en prijsinformatie voorbehouden.

Normen als digitaal document? Bestel online via [www.nen-normshop.nl](http://www.nen-normshop.nl)